

Calibration Certificate



Calibration Performed By:

EVEREST SCALE, INC.
14 DUNBAR STREET
GREENVILLE, SC 29601
864-242-5885

For:

ABC COMPANY
123 MAIN STREET
YOURTOWN, SC 29000
JOHN ADAMS

Equipment Information

I.D.:	FS-12	Performed By:	SCHINTZIUS	Cal Date:	3/7/2016
Description:	FLOOR SCALE	Test Location:	STAPLE	Cal. Due Date:	6/7/2016
Manufacturer:	METTLER TOLEDO	Model Number:		Cal. Interval:	3 MONTHS
Gage Type:	FLOOR SCALE	Serial Number:	8704-255	Calibration Result:	ADJ.
Temp./RH:	0 F / 0	Capacity/Graduation:	2000 LB / .2 LB		
PO Number:	123654				

Calibration Notes

Test Points

Seq.	Description	Standard	Tolerance -	Tolerance +	As Found	As Left
1	Zero Checked	0.0 lb	-0.2	0.2	0.0	0.0
2	Shift Test Quadrant 1	500.0 lb	499.4	500.6	500.0	500.0
3	Shift Test Quadrant 2	500.0 lb	499.4	500.6 A	496.0	500.0
4	Shift Test Quadrant 3	500.0 lb	499.4	500.6	500.0	500.0
5	Shift Test Quadrant 4	500.0 lb	499.4	500.6 A	499.0	500.0
6	Zero Checked	0.0 lb	-0.2	0.2	0.0	0.0
7	Increasing Load	500.0 lb	499.4	500.6	500.0	500.0
8	Increasing Load	750.0 lb	749.4	750.6 A	749.0	750.0
9	Increasing Load	1000.0 lb	999.0	1001.0	1001.0	1000.0
10	Decreasing Load	750.0 lb	749.4	750.6	749.4	750.0
11	Decreasing Load	500.0 lb	499.4	500.6	499.4	500.0
12	Zero Checked	0.0 lb	-0.2	0.2	0.0	0.0

Standards Used To Calibrate Equipment

Weight ID #	State Test Number	Last Cal.	Cal. Due Date
TRK9-1	SC730529-20M	7/9/2015	7/9/2016
TRK9-10	SC820505-15M	7/9/2015	7/9/2016
TRK9-2	SC730529-21M	7/9/2015	7/9/2016
TRK9-3	SC730529-18M	7/9/2015	7/9/2016
TRK9-4	SC730529-91M	7/9/2015	7/9/2016
TRK9-5	SC730529-87M	7/9/2015	7/9/2016
TRK9-6	SC730529-123M	7/9/2015	7/9/2016
TRK9-7	SC900209-114M	7/9/2015	7/9/2016
TRK9-8	SC730529-23M	7/9/2015	7/9/2016
TRK9-9	SC730529-72M	7/9/2015	7/9/2016

Calibration Certificate



Certified By:

ROBERT SCHINTZIUS

Accepted by:

CJS

Registration No.: 202508

Date: 3/7/2016

Standards used are traceable to the National Institute of Standards and Technology (NIST).

Certificate reproduction except in full requires the written permission from Everest Scale, Inc. This document may contain confidential or privileged information and is intended for the individual or entity named above. If you are not the intended recipient, be aware that any disclosure, copying, distribution or use of the contents of this information is prohibited. If you have received this document in error, please contact Everest Scale, Inc. immediately at (864) 242-5885.